

RAZORBLADE

TATTOO, PIERCING, AND MEDICAL SUPPLIES

INSPECTION CERTIFICATE / CERTIFICATE OF CONFORMITY

ABNAHMEPRÜFZEUGNIS CERTIFICAT DE RÉCEPTION

PRODUCT ERZEUGNISFORM PRODUIT	316L Stainless Steel							
SPECIFICATION LIEFERBEDINGUNGEN CONDITION DE LIVRAISON	316S11 (316L Stainless Steel)							
CHEMICAL COMPOSITION CHEMISCHE ZUSAMMENSETZUNG COMPOSITION CHIMIQUE	C 0.011	Si 0.33	Mn 1.39	S 0.001	P 0.023	Cr 19.91	Ni 11.51	Mo 2.20
REMARKS BEMERKUNGEN REMARQUES	Complies With European Directive 94/27/EC for use after epithelisation Nickel Content Test In Accordance With BS EN 1810 – 1998 Nickel Release Test In Accordance With BS EN 1811 – 1999							

PRODUCT ERZEUGNISFORM PRODUIT	Implant Titanium							
SPECIFICATION LIEFERBEDINGUNGEN CONDITION DE LIVRAISON	6Al 4V ELi to BS7252 For Implant Use							
CHEMICAL COMPOSITION CHEMISCHE ZUSAMMENSETZUNG COMPOSITION CHIMIQUE	C 0.008	N 0.009	Fe 0.18	Al 6.25	V 4.19	O 0.19	Ti Bal	
REMARKS BEMERKUNGEN REMARQUES	Complies Fully With European Directive 94/27/EC Nickel Content Test In Accordance With BS EN 1810 – 1998 Nickel Release Test In Accordance With BS EN 1811 – 1999							

Razorblade Pro
22 N. Cameron Street
Harrisburg, PA 17101
(877) 777 - 2967
www.razorblade.pro

